


Mass Spectrum



Sample Information

Month-Day Processed : 05/12/21

Time Processed : 7:31:48 AM

Injection Volume : 0.3

Sample Name : Influenza NP (147-155)

Cat. No. : RP20260

Lot. No. : E1981621E

Theoretical MW : 1107.27

Observed MW : 1107.0

Equipment : GK11010007

Interface Bias : +4.5 kV

Drying Gas Flow : 5 L/min

T.Flow : 0.2 ml/min

B.conc : 50% H2O/50% MeOH

Interface : ESI

Nebulizing Gas Flow : 1.5 L/min

CDL Temp : 250

Block Temp : 200