

Sample Name :CTTHWGFTLC, CYCLIC
 Cat No. :RP20386
 Lot No. :PE8811905
 Time Processed :1:23:27
 Month-Day-Year Processed :05/12/19

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

| Time | Module | Command | Value |
|-------|------------|-----------------|-------|
| 0.01 | Pumps | Solvent B Conc. | 5 |
| 25.00 | Pumps | Solvent B Conc. | 65 |
| 25.01 | Pumps | Solvent B Conc. | 95 |
| 27.00 | Pumps | Solvent B Conc. | 95 |
| 27.01 | Pumps | Solvent B Conc. | 5 |
| 33.00 | Pumps | Solvent B Conc. | 5 |
| 33.01 | Controller | Stop | |


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm
 Equipment: ZJ17010508

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

| Peak# | Ret. Time | Area | Height | Area% |
|-------|-----------|----------|---------|---------|
| 1 | 15.528 | 21055 | 1775 | 0.167 |
| 2 | 15.764 | 7150 | 862 | 0.057 |
| 3 | 15.996 | 3446 | 473 | 0.027 |
| 4 | 16.364 | 9669 | 996 | 0.077 |
| 5 | 16.903 | 156414 | 12826 | 1.241 |
| 6 | 17.209 | 45781 | 5233 | 0.363 |
| 7 | 17.563 | 12352927 | 1134466 | 98.005 |
| 8 | 18.008 | 1173 | 238 | 0.009 |
| 9 | 18.232 | 2980 | 366 | 0.024 |
| 10 | 23.609 | 1385 | 118 | 0.011 |
| 11 | 24.894 | 2375 | 251 | 0.019 |
| Total | | 12604355 | 1157605 | 100.000 |