

Sample Name: CEF3
 Cat No. : RP11765
 Lot No. : E4020421F
 Time Processed : 1:20:18 PM
 Month-Day-Year Processed : 06/16/2021

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow: 1 ml/min
 Wavelength: 220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Pump A B.Conc	5
25.00	Pumps	Pump A B.Conc	65
25.01	Pumps	Pump A B.Conc	95
27.00	Pumps	Pump A B.Conc	95
27.01	Pumps	Pump A B.Conc	5
35.00	Pumps	Pump A B.Conc	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm
 Equipment: ZJ19010141

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	9.509	1175	166	0.051
2	10.407	1585	196	0.068
3	11.427	1000	140	0.043
4	14.265	1878	161	0.081
5	15.171	11343	1382	0.489
6	15.715	2669	329	0.115
7	15.917	2250337	313556	96.960
8	16.291	3026	488	0.130
9	16.723	4238	562	0.183
10	17.019	2693	435	0.116
11	17.381	32647	4150	1.407
12	19.724	2992	415	0.129
13	23.512	1388	233	0.060
14	23.757	3921	440	0.169

Peak#	Ret. Time	Area	Height	Area%
Total		2320893	322653	100.000