

Sample Name : OVA Peptide (257-264)
 Cat. No. : RP10611
 Lot. No. : PE4841610

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm


Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
25.00	Pumps	Pump A B.Conc	65	
25.01	Pumps	Pump A B.Conc	95	
31.00	Pumps	Pump A B.Conc	95	
31.01	Pumps	Pump A B.Conc	5	
40.00	Pumps	Pump A B.Conc	5	
40.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : Alltima™ C18 4.6 x 250 mm

Chromatogram



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	15.591	4230	377	0.224
2	16.159	2269	328	0.120
3	17.397	7388	1220	0.391
4	17.558	6654	1256	0.352
5	17.754	13166	1944	0.697
6	17.951	1847877	290069	97.808
7	18.649	2610	484	0.138
8	21.042	5092	831	0.270
Total		1889286	296508	100.000