

Sample Name :TAT  
 Cat. No. : RP12784  
 Lot. No. : PE2741806


Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm

Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
25.00	Pumps	Pump A B.Conc	65	
25.01	Pumps	Pump A B.Conc	95	
27.00	Pumps	Pump A B.Conc	95	
27.01	Pumps	Pump A B.Conc	5	
35.00	Pumps	Pump A B.Conc	5	
35.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm



1 Det.A Ch1 / 220nm  
 Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	6.384	21505	3737	0.949
2	7.229	7509	1435	0.332
3	8.419	10749	2038	0.475
4	8.953	9619	1824	0.425
5	9.318	25884	3466	1.143
6	9.559	2159572	321198	95.347
7	9.868	27254	4817	1.203
8	16.359	2873	423	0.127
Total		2264962	338939	100.000