

Sample Name :InsB 9-23
 Cat No.: RP20391
 Lot No.: E6045320C
 Time Processed :17:52:59
 Month-Day-Year Processed :03/27/2020

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>


Time	Module	Command	Value
0.01	Pumps	Solvent B Conc.	5
25.00	Pumps	Solvent B Conc.	65
25.01	Pumps	Solvent B Conc.	95
27.00	Pumps	Solvent B Conc.	95
27.01	Pumps	Solvent B Conc.	5
35.00	Pumps	Solvent B Conc.	5
35.01	Controller	Stop	

<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm
 Equipment:ZJ19010014

<Chromatogram>



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	13.932	2677	539	0.016
2	14.351	5363	775	0.032
3	14.559	2602	516	0.016
4	14.704	2493	318	0.015
5	15.156	2678	283	0.016
6	15.368	51733	6383	0.312
7	15.873	18396	3286	0.111
8	16.164	4503	672	0.027
9	16.299	1915	389	0.012
10	16.601	29344	4494	0.177
11	16.810	122863	11836	0.740
12	17.179	15933642	1912797	95.961
13	17.579	383147	31749	2.308
14	17.893	34620	4928	0.209
15	18.247	4780	832	0.029

Peak#	Ret. Time	Area	Height	Area%
16	18.601	2093	400	0.013
17	18.813	1363	277	0.008
Total		16604212	1980473	100.000