

Sample Name :Bradykinin
 Cat. No. : RP11129
 Lot. No. : PE7011710

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm


Time	Unit	Command	Value	Comment
0.01	Pumps	Pump B Conc.	5	
25.00	Pumps	Pump B Conc.	65	
25.01	Pumps	Pump B Conc.	95	
31.00	Pumps	Pump B Conc.	95	
31.01	Pumps	Pump B Conc.	5	
40.00	Pumps	Pump B Conc.	5	
40.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : Alltima™ C18 4.6 x 250 mm

Chromatogram



I Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	8.407	27131	3246	0.230
2	9.235	8560	1380	0.073
3	10.989	1990	397	0.017
4	11.194	3160	570	0.027
5	11.612	20189	2974	0.171
6	12.094	10407	1705	0.088
7	12.336	1764	383	0.015
8	13.597	33108	6860	0.281
9	13.764	10126	1589	0.086
10	14.056	9545	2329	0.081
11	14.229	390600	63914	3.317
12	14.504	11198455	1257220	95.107
13	14.787	59576	15182	0.506
Total		11774611	1357749	100.000