

Sample Name :Axl tide  
 Cat. No. : RP20363  
 Lot. No. : E8005920I  
 Time Processed: 11:01:42 AM  
 Month-Day-Year Processed :10/4/2020

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Pump A B.Conc	5
25.00	Pumps	Pump A B.Conc	65
25.01	Pumps	Pump A B.Conc	95
31.00	Pumps	Pump A B.Conc	95
31.01	Pumps	Pump A B.Conc	5
40.00	Pumps	Pump A B.Conc	5
41.00	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm  
 Equipment:ZJ19010325

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	8.502	130515	19463	0.702
2	8.605	136695	19085	0.736
3	9.072	17836551	1571268	96.003
4	9.419	121110	15349	0.652
5	9.525	27652	4845	0.149
6	9.831	301130	42122	1.621
7	10.710	25478	4722	0.137
Total		18579132	1676854	100.000