

Sample Name :GRGDSP
 Cat No. :RP20255
 Lot No. :E7238221A
 Time Processed :14:00:17
 Month-Day-Year Processed :01/19/2021

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	B.Conc	0
25.00	Pumps	B.Conc	50
25.01	Pumps	B.Conc	95
27.00	Pumps	B.Conc	95
27.01	Pumps	B.Conc	5
35.00	Pumps	B.Conc	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm
 Equipment: ZJ19010324

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	7.818	25867	4834	0.557
2	8.487	29191	1810	0.628
3	8.682	4585870	848736	98.690
4	9.149	1225	182	0.026
5	9.300	3083	308	0.066
6	9.581	1530	340	0.033
Total		4646765	856210	100.000