

Sample Name:MOG(35-55)
 Cat No.: RP10245
 Lot No.: E2633720A
 Time Processed :4:08:05
 Month-Day-Year Processed :01/15/2020

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Solvent B Conc.	5
25.00	Pumps	Solvent B Conc.	65
25.01	Pumps	Solvent B Conc.	95
27.00	Pumps	Solvent B Conc.	95
27.01	Pumps	Solvent B Conc.	5
33.00	Pumps	Solvent B Conc.	5
33.00	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm
 Equipment:ZJ19010324

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	9.158	1594	183	0.017
2	9.409	20251	525	0.218
3	10.533	2571	283	0.028
4	10.739	13293	597	0.143
5	11.671	6525	258	0.070
6	12.627	27259	1046	0.293
7	13.431	3449	320	0.037
8	13.710	4339	778	0.047
9	13.877	6941	883	0.075
10	14.136	99554	11252	1.071
11	14.315	12538	2362	0.135
12	14.543	9027571	1470374	97.160
13	15.063	5245	980	0.056
14	15.788	7860	129	0.085
15	16.852	23620	1935	0.254

Peak#	Ret. Time	Area	Height	Area%
16	17.938	3928	248	0.042
17	18.309	3284	178	0.035
18	18.907	7109	738	0.077
19	19.257	1109	137	0.012
20	19.723	1242	109	0.013
21	20.237	1120	114	0.012
22	20.703	1461	172	0.016
23	22.060	8210	852	0.088
24	22.584	1408	121	0.015
Total		9291482	1494575	100.000