

Sample Name :NP (396-404)
 Cat. No. : RP20090
 Lot. No. : PE2011905
 Time Processed :10:44:29 AM
 Month-Day-Year Processed :06/07/2019

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Solvent B Conc.	5
25.00	Pumps	Solvent B Conc.	65
25.01	Pumps	Solvent B Conc.	95
27.00	Pumps	Solvent B Conc.	95
27.01	Pumps	Solvent B Conc.	5
35.00	Pumps	Solvent B Conc.	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm
 Equipment:ZJ19010141

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	12.917	5033437	822885	96.385
2	13.131	10799	2843	0.207
3	13.451	33526	6478	0.642
4	13.673	8865	1646	0.170
5	14.139	1742	422	0.033
6	14.340	13314	2052	0.255
7	15.290	44847	7195	0.859
8	15.424	75682	9509	1.449
Total		5222212	853029	100.000