

Sample Name :RGD  
 Cat. No. : RP20297  
 Lot. No. : PE2331712


Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm

Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
25.00	Pumps	Pump A B.Conc	65	
25.01	Pumps	Pump A B.Conc	95	
27.00	Pumps	Pump A B.Conc	95	
27.01	Pumps	Pump A B.Conc	5	
35.00	Pumps	Pump A B.Conc	5	
35.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm



Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	11.476	7050	1216	0.061
2	11.934	81957	9673	0.710
3	12.266	78914	14588	0.684
4	12.449	11147784	1546061	96.639
5	12.718	164668	23845	1.427
6	12.937	55096	5533	0.478
Total		11535469	1600916	100.000