

Sample Name :N-Cbz-Phe-Arg-AMC
 Cat No.: RP11385
 Lot No.: E1308020E
 Time Processed :8:18:21
 Month-Day-Year Processed :05/15/2020

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Solvent B Conc.	5
25.00	Pumps	Solvent B Conc.	65
25.01	Pumps	Solvent B Conc.	95
27.00	Pumps	Solvent B Conc.	95
27.01	Pumps	Solvent B Conc.	5
35.00	Pumps	Solvent B Conc.	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm
 Equipment:ZJ19010014

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	14.328	1475	261	0.002
2	14.456	3236	638	0.005
3	14.672	20843	3119	0.034
4	14.856	67368	12517	0.111
5	19.978	1637	294	0.003
6	20.162	17148	1304	0.028
7	21.315	2781	447	0.005
8	21.588	29540	3219	0.049
9	21.876	5735	788	0.009
10	22.212	60269963	2560371	99.356
11	22.898	74167	8938	0.122
12	23.147	5078	847	0.008
13	23.280	4158	697	0.007
14	23.502	18674	3100	0.031
15	23.767	2339	466	0.004

Peak#	Ret. Time	Area	Height	Area%
16	24.298	88932	10761	0.147
17	24.726	45446	7365	0.075
18	24.969	2238	506	0.004
Total		60660758	2615637	100.000