

Sample Name :InsB 9-23
 Cat No. :RP20391
 Lot No. :E7173820L
 Time Processed :11:53:48
 Month-Day-Year Processed :01/05/2021

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm
 <<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	B.Conc	5
25.00	Pumps	B.Conc	65
25.01	Pumps	B.Conc	95
27.00	Pumps	B.Conc	95
27.01	Pumps	B.Conc	5
35.00	Pumps	B.Conc	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm
 Equipment: ZJ19010141

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	12.630	15555	2428	0.223
2	12.857	65950	10592	0.947
3	13.072	21797	3412	0.313
4	13.277	41579	4087	0.597
5	13.665	6617629	910260	95.003
6	14.007	91717	12807	1.317
7	14.193	7570	1193	0.109
8	14.449	18958	2961	0.272
9	14.629	18068	2853	0.259
10	14.774	9167	1562	0.132
11	15.396	6954	1161	0.100
12	15.994	13409	2030	0.192
13	17.311	18821	2997	0.270
14	23.292	18566	2457	0.267

Peak#	Ret. Time	Area	Height	Area%
Total		6965739	960800	100.000