

Sample Name :Adrenomedullin (AM) (13-52), human
 Cat No. :RP11291
 Lot No. :E9409920B
 Time Processed :10:14:48
 Month-Day-Year Processed :02/08/2020

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Solvent B Conc.	5
25.00	Pumps	Solvent B Conc.	65
25.01	Pumps	Solvent B Conc.	95
27.00	Pumps	Solvent B Conc.	95
27.01	Pumps	Solvent B Conc.	5
35.00	Pumps	Solvent B Conc.	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Alltima™ C18 4.6 x 250 mm
 Equipment:ZJ19010141

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	12.054	11048	560	0.245
2	12.678	6137	968	0.136
3	13.181	4495260	374689	99.567
4	13.992	1217	228	0.027
5	14.206	1148	127	0.025
Total		4514811	376573	100.000