

Sample Name : Arg-Gly-Asp-Ser
 Cat. No. : RP10861
 Lot. No. : E2025321C
 Time Processed : 6:07:09 AM
 Month-Day-Year Processed : 04/10/2021
 Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow: 1 ml/min
 Wavelength: 220 nm
 <<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Pump A B.Conc	0
5.00	Pumps	Pump A B.Conc	0
20.00	Pumps	Pump A B.Conc	30
25.00	Pumps	Pump A B.Conc	65
25.01	Pumps	Pump A B.Conc	95
27.00	Pumps	Pump A B.Conc	95
27.01	Pumps	Pump A B.Conc	0
35.00	Pumps	Pump A B.Conc	0
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm
 Equipment : SS-CM-0286

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	3.784	3287575	426694	99.723
2	4.268	1744	324	0.053
3	5.085	314	65	0.010
4	5.294	502	62	0.015
5	5.375	218	57	0.007
6	5.452	178	57	0.005
7	5.522	323	57	0.010
8	5.712	587	76	0.018
9	5.917	342	67	0.010
10	6.040	717	92	0.022
11	6.226	786	117	0.024
12	6.308	183	64	0.006
13	6.392	266	63	0.008
14	6.474	532	86	0.016
15	6.608	118	38	0.004

Peak#	Ret. Time	Area	Height	Area%
16	6.721	331	66	0.010
17	6.804	468	94	0.014
18	6.967	444	67	0.013
19	7.119	405	79	0.012
20	7.250	210	44	0.006
21	7.320	242	73	0.007
22	7.482	233	40	0.007
Total		3296715	428382	100.000