

Sample Name : CEF4  
 Cat No.: RP11766  
 Lot No.: E4023621F  
 Time Processed : 13:35:55  
 Month-Day-Year Processed : 06/16/2021

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow: 1 ml/min  
 Wavelength: 220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Solvent B Conc.	5
25.00	Pumps	Solvent B Conc.	65
25.01	Pumps	Solvent B Conc.	95
27.00	Pumps	Solvent B Conc.	95
27.01	Pumps	Solvent B Conc.	5
35.00	Pumps	Solvent B Conc.	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm  
 Equipment: ZJ19010014

### <Chromatogram>

mV



### <Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	11.682	1176	196	0.172
2	12.209	649749	92407	95.087
3	12.880	15694	1910	2.297
4	13.351	9076	1181	1.328
5	13.502	3291	519	0.482
6	18.763	4332	569	0.634
Total		683317	96783	100.000