

Sample Information

Sample Name :InsB 9-23
 Cat No :RP20391
 Lot No :PE1631808

Pump A : 0.065% trifluoroacetic in 100% waters (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)


<<Pump>>
 Flow : 1.0000 mL/min

<<Detector A>>
 Wavelength Ch1 : 220 nm

<<LC Time Program>>

Time	Module	Command	Value	Comment
0.01	Pumps	Solvent B Conc.	5	
25.00	Pumps	Solvent B Conc.	65	
25.01	Pumps	Solvent B Conc.	95	
31.00	Pumps	Solvent B Conc.	95	
31.01	Pumps	Solvent B Conc.	5	
40.00	Pumps	Solvent B Conc.	5	
40.01	Controller	Stop		

<<Column Performance>>
 <Detector A>
 Column : Inertsil ODS-3 4.6 x 250 mm



Peak Table

Peak#	Ret. Time	Height	Area	Height%	Area%
1	12.329	7675	37075	1.716	1.264
2	12.897	2406	13929	0.538	0.475
3	13.319	423077	2792987	94.584	95.232
4	13.682	2389	7590	0.534	0.259
5	13.851	2768	15896	0.619	0.542
6	16.545	6668	40577	1.491	1.384
7	20.767	2320	24761	0.519	0.844
Total		447303	2932813	100.000	100.000