

Sample Name : HA-(126-138)
 Cat. No. : RP20240
 Lot. No. : CC05908


Pump A : 0.065% trifluoroacetic in 100% water
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile
 Total Flow : 1 ml / min
 Wavelength : 220 nm

Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
25.00	Pumps	Pump A B.Conc	65	
25.01	Pumps	Pump A B.Conc	95	
35.00	Pumps	Pump A B.Conc	95	
35.01	Pumps	Pump A B.Conc	5	
45.00	Pumps	Pump A B.Conc	5	
45.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : AlltimaTM C18 4.6 x 250 mm



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	10.068	20306	3206	0.377
2	10.177	17719	2900	0.329
3	10.468	5151828	848642	95.746
4	10.861	12308	1912	0.229
5	11.658	94227	12955	1.751
6	17.817	55263	9654	1.027
7	21.012	29098	4155	0.541
Total		5380750	883423	100.000