

Sample Name : G3-C12
Cat. No. :RP20329
Lot. No. :CB04910


Pump A : 0.065% trifluoroacetic in 100% water
Pump B : 0.05% trifluoroacetic in 100% acetonitrile
Total Flow:1 ml/min
Wavelength:220 nm

Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
25.00	Pumps	Pump A B.Conc	65	
25.01	Pumps	Pump A B.Conc	95	
35.00	Pumps	Pump A B.Conc	95	
35.01	Pumps	Pump A B.Conc	5	
45.00	Pumps	Pump A B.Conc	5	
45.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : Alltima™ C18 4.6 x 250 mm



Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	8.616	3219	773	0.068
2	10.677	1416	278	0.030
3	11.066	2442	378	0.052
4	11.282	31330	3496	0.667
5	11.489	7337	1210	0.156
6	11.750	47786	7109	1.017
7	11.918	4477646	752005	95.278
8	12.161	104966	9278	2.234
9	12.613	19051	2882	0.405
10	13.165	1212	276	0.026
11	13.902	3173	649	0.068
Total		4699578	778336	100.000