

Sample Name : CEF27
 Cat. No. :RP20397
 Lot. No. :84280007112509LQ


Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
25.00	Pumps	Pump A B.Conc	65	
25.01	Pumps	Pump A B.Conc	95	
35.00	Pumps	Pump A B.Conc	95	
35.01	Pumps	Pump A B.Conc	5	
45.00	Pumps	Pump A B.Conc	5	
45.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : AlltimaTM C18 4.6 x 250 mm



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	9.206	20507	3604	0.184
2	9.485	9852	2002	0.088
3	9.596	10646	2029	0.096
4	9.759	22236	5141	0.200
5	9.843	37764	7428	0.339
6	9.937	27460	5036	0.246
7	10.150	59011	13193	0.529
8	10.260	10794322	1770323	96.855
9	10.584	42822	8205	0.384
10	10.669	73291	10265	0.658
11	10.943	9600	807	0.086
12	11.286	14792	1159	0.133
13	11.729	7299	1491	0.065
14	12.942	15200	2671	0.136
Total		11144804	1833354	100.000