

CERTIFICATE OF ANALYSIS

Product Information

Product Name	PEGylated Molecule Assay Kit
Cat. No.	L00458
Lot No.	16E001045
Kit Content	The kit contains 11 components PEG Molecule Plate, Assay Diluent, Biotin-PEG mAb Stock Antibody Dilution Buffer, Streptavidin-HRP, PEG Standard Control, 20 x Wash Solution, TMB Substrate, Stop Solution, Plate Sealer And User Manual
Storage	The product is stable for at least one year if stored at 2-8 °C

Quality Control

Test Items	Specifications	Results
IC50	<1 ng/ml for PEG40 kDa	Passed
Range	0.333~81 ng/ml	Passed

Caution

For laboratory or further manufacturing use only. Not intended for household use. If you have any questions about the Certificate of Analysis, please contact our customer service representative at 1-877-436-7274 (Toll-Free), or 1-732-885-9188.

Certified by: *Higgin chin*

Antibody Production Manager

Date: 05/12/2016