

CERTIFICATE OF ANALYSIS

Product Information

Product Name	Human IL-8 ELISA Detection Kit
Cat. No.	L00417
Lot No.	16J001121
Kit Content	The kit contains 10 components IL-8 Capture Plate, IL-8 Detection Reagent, IL-8 Standard, Sample Diluent, Assay Diluent, 20 × Wash Solution, TMB Substrate, Stop Solution, Adhesive Plate Cover, Technical Manual
Storage	The product is stable for at least one year if stored at 2-8 °C

Quality Control

Test Items	Specifications	Results
Minimum Detectable Dose	<1 ng/ml	Passed
Dynamic Range	25~1600 pg/ml	Passed
R ²	>99%	Qualified
Zero OD ₄₅₀ Absorbance Value	<0.1 unit	Qualified

Caution

For laboratory or further manufacturing use only. Not intended for household use. If you have any questions about the Certificate of Analysis, please contact our customer service representative at 1-877-436-7274 (Toll-Free), or 1-732-885-9188.

Certified by: 
Antibody Production Director Date: 10/21/2016