

CERTIFICATE OF ANALYSIS

Product Information

| | |
|-----------|---|
| Name | cGMP ELISA Detection Kit |
| Cat. No. | L00461 |
| Lot No. | 17F001269 |
| Component | Anti-IgG Captured Plate Anti-cGMP pAb HRP-cGMP cGMP Standards (0, 0.3, 0.8, 2.5, 7.4, 22.2, 66.7 pmol/ml) cGMP Standard Stock (10 nmol/ml) Assay Buffer A Assay Buffer B 20 × Wash Solution Stop Solution TMB Substrate Plate Sealer User Manual |
| Storage | The unopened kit is stable for at least 12 months if stored at 2-8 °C and the opened kit is stable for up to 2 weeks at 2-8 °C. Do not freeze the kit. |


Quality Control

| Test Items | Specifications | Results |
|-----------------------|----------------------|-----------|
| Reactivity | Can detect cGMP | Conformed |
| Lower Detection Limit | 0.11 pmol/ml of cGMP | Qualified |
| Dynamic Range | 0.3-66.7 pmol/ml | Qualified |

Caution

For laboratory or further manufacturing use only. Not intended for household use. If you have any questions about the Certificate of Analysis, please contact our customer service representative at 1-877-436-7274 (Toll-Free), 1-732-885-9188.

Certified by:



Date: 06/28/2017

Antibody Production Director