

## **CERTIFICATE OF ANALYSIS**

## **Product Information**

| i roddot iiiioriiiatioii |  |
|--------------------------|--|
| Product Name             | His Tag ELISA Detection Kit                            |
| Cat. No.                 | L00436   |
| Lot No.                  | 18B001426  |
| Kit Content              | His Tag Plate (8 wells x 12 strips)                    |
|                          | Anti-His Monoclonal Antibody                           |
|                          | Antibody Tracer  |
|                          | His Tag Standards (0, 1, 3, 9, 27, 81, 243, 729 ng/ml) |
|                          | Assay Diluent  |
|                          | 20 x Wash Solution                                     |
|                          | TMB Substrate  |
|                          | Stop Solution  |
|                          | His Tag Stardard Stock (10 μg/ml)                      |
|                          | Plate Sealer   |
|                          | User Manual  |
| Storage                  | Store at 2-8 °C  |

## **Quality Control**

| Test Items            | Specifications                | Results   |
|-----------------------|-------------------------------|-----------|
| Reactivity            | Can detect His-tagged protein | Passed    |
| Lower Detection Limit | 1 ng/ml                       | Qualified |
| Dynamic Range         | 1~729 ng/ml                   | Qualified |

## Caution

For laboratory or further manufacturing use only. Not intended for household use. If you have any questions about the Certificate of Analysis, please contact our customer service representative at 1-877-436-7274 (Toll-Free), or 1-732-885-9188.

Certified by:

Higgins Chin Date: 02/28/2018

**Antibody Production Director**