

CERTIFICATE OF ANALYSIS

Product Information

| | |
|--------------|---|
| Product Name | PEGylated Molecule Assay Kit |
| Cat. No. | L00458 |
| Lot No. | 18D001462 |
| Kit Content | The kit contains 11 components PEG Molecule Plate, Assay Diluent, Biotin-PEG mAb Stock Antibody Dilution Buffer, Streptavidin-HRP, PEG Standard Control, 20 × Wash Solution, TMB Substrate, Stop Solution, Plate Sealer And User Manual |
| Storage | The product is stable for at least one year if stored at 2-8 °C |

Quality Control

| Test | Items | Specifications | Results |
|-------|-------|------------------------|---------|
| IC50 | | <1 ng/ml for PEG40 kDa | Passed |
| Range | | 0.333~81 ng/ml | Passed |

Caution

For laboratory or further manufacturing use only. Not intended for household use. If you have any questions about the Certificate of Analysis, please contact our customer service representative at 1-877-436-7274 (Toll-Free), or 1-732-885-9188.

Certified by: 
Antibody Production Manager

Date: 04/26/2018