


Sample Description:
 Structure : RP10354
 Column : 4.6×250mm, Sinochrom ODS-BP-5
 Solvent A : 0.1% trifluoroacetic in 100% acetonitrile
 Solvent B : 0.1% trifluoroacetic in 100% water
 Gradient :
 A B
 0.01min 15% 85%
 25min 55% 45%
 25.1min 100% 0%
 30min STOP
 Flow rate : 1.0 mL/min
 Wavelength : 220nm

Cat. No.: RP10354
 Lot No.: PY00326507



Results

Peak No.	Peak ID	Ret Time	Height	Area	Conc.
1		13.050	188470.063	1389031.250	95.1689
2		13.453	7238.489	44111.336	3.0223
3		13.612	1286.500	4832.388	0.3311
4		13.995	3881.920	21568.756	1.4778
Total			200876.972	1459543.729	100.0000