

CERTIFICATE OF ANALYSIS

Product Information

Name	cGMP ELISA Detection Kit
Cat. No.	L00461
Lot No.	19D001777
Component	Anti-IgG Captured Plate Anti-cGMP pAb HRP-cGMP cGMP Standards (0, 0.3, 0.8, 2.5, 7.4, 22.2, 66.7 pmol/ml) cGMP Standard Stock (10 nmol/ml) Assay Buffer A Assay Buffer B 20 × Wash Solution Stop Solution TMB Substrate Plate Sealer User Manual
Storage	The unopened kit is stable for at least 12 months if stored at 2-8 °C and the opened kit is stable for up to 2 weeks at 2-8 °C. Do not freeze the kit.

Quality Control

Test Items	Specifications	Results
Reactivity	Can detect cGMP	Conformed
Lower Detection Limit	0.11 pmol/ml of cGMP	Qualified
Dynamic Range	0.3-66.7 pmol/ml	Qualified

Caution

For laboratory or further manufacturing use only. Not intended for household use. If you have any questions about the Certificate of Analysis, please contact our customer service representative at 1-877-436-7274 (Toll-Free), 1-732-885-9188.

Certified by:  Date: 05/06/2019

Antibody Production Director