

CERTIFICATE OF ANALYSIS

Product Information

| | |
|--------------|---|
| Product Name | His Tag ELISA Detection Kit |
| Cat. No. | L00436 |
| Lot No. | 19F001813 |
| Kit Content | His Tag Plate (8 wells x 12 strips) Anti-His Monoclonal Antibody Antibody Tracer His Tag Standards (0, 1, 3, 9, 27, 81, 243, 729 ng/ml) Assay Diluent 20 x Wash Solution TMB Substrate Stop Solution His Tag Standard Stock (10 µg/ml) Plate Sealer User Manual |
| Storage | Store at 2-8 °C |

Quality Control

| Test Items | Specifications | Results |
|-----------------------|-------------------------------|-----------|
| Reactivity | Can detect His-tagged protein | Passed |
| Lower Detection Limit | 1 ng/ml | Qualified |
| Dynamic Range | 1~729 ng/ml | Qualified |

Caution

For laboratory or further manufacturing use only. Not intended for household use. If you have any questions about the Certificate of Analysis, please contact our customer service representative at 1-877-436-7274 (Toll-Free), or 1-732-885-9188.

Certified by:  Date: 06/13/2019
Antibody Production Director