

Summary:

Structure : GG-10
Column : 4.6×250mm, Kromasil C18-5
Solvent A : 0.1% trifluoroacetic in 100% acetonitrile
Solvent B : 0.1% trifluoroacetic in 100% water
Gradient :


	A	B
0.01min	17%	83%
25min	40%	60%
25.1min	100%	0%
30min	STOP	

Flow rate : 1.0 mL/min

Wavelength : 220nm

Cat. No.: RP10620

Lot No.: PJ00610505



RESULTS

Peak No.	Peak ID	Ret Time	Height	Area	Conc
1		13.138	117.244	844.550	0.1779
2		13.675	90.381	474.750	0.1000
3		13.897	112.124	685.949	0.1445
4		14.480	5011.535	456878.688	96.2587
5		14.893	1837.250	15752.504	3.3189
TOTAL			5248.535	474636.440	100.0000