


Sample Description:
 Structure : YM-5-NH2
 Column : 4.6×250mm, VYDAC-C18
 Solvent A : 0.1% trifluoroacetic in 100% acetonitrile
 Solvent B : 0.1% trifluoroacetic in 100% water
 Gradient :
 A B
 0.01min 15% 85%
 25min 45% 55%
 25.1min 100% 0%
 30min STOP
 Flow rate : 1.0 mL/min
 Wavelength : 220nm

Cat. No.: RP10887
 Lot No.: PX00211503



Results

Peak No.	Peak ID	Ret Time	Height	Area	Conc.
1		6.863	2178.156	36834.160	0.0958
2		7.643	931088.375	38335824.000	99.7249
3		9.172	4293.769	53727.742	0.1398
4		10.868	574.951	10884.704	0.0283
5		11.363	321.897	4323.865	0.0112
Total			938457.148	38441594.471	100.0000