

Version 1.1

Update: 12/27/2019

CERTIFICATE OF ANALYSIS

Product Information

Product Name	Protein A ELISA Kit		
Cat. No.	L00430		
Lot No.	2103K005		
Kit Content	The kit contains 13 components:		
	96-well plate (8 wells x 12 strips) coated with anti-Protein A mAb, Biotin		
	conjugated anti-Protein A mAb, Streptavidin-HRP, Recombinant Protein A		
	Standard (10 μg/ml),		
	MabSelect SuRe™ Protein A Standard (10 μg/ml), Antibody Dilution Buffer,		
	20 x Wash Solution, 5 x Sample dilution Solution, 10% Tween 20, One-		
	Solution Microwell TMB substrate, Stop Solution, Plate sealers and User		
	Manual		
Storage	The product is stable for at least one year if stored at 2-8 °C		

Quality Control

Test Items	Specifications	Results
Reactivity	The kit can detect recombinant Protein A	Passed
	variants, native Protein A and MabSelect SuRe™	
	Protein A in the presence or absence of IgG	
	antibodies.	
Lower Detection Limit	20 pg/ml of Recombinant Protein A	Qualified
	50 pg/ml of MabSelect SuRe™ Protein A	
Dynamic Range:	0.00-1.6 ng/ml (Recombinant Protein A)	Qualified
	0.00-3.2 ng/ml (MabSelect SuRe™ Protein A)	
Intra-assay Precision CV	<5%	Qualified
Inter-assay Precision CV	<10%	Qualified
Recovery	Within allowable limits 85% to 115%	Qualified

Caution

For laboratory or further manufacturing use only. Not intended for household use. If you have any questions about the Certificate of Analysis, please contact our customer service representative at 1-877-436-7274 (Toll-Free), or 1-732-885-9188.

Certified by: Xianli Yang Date: 03/29/2021

Quality Department

For laboratory research use only. Direct human use, including taking orally and injection and clinical use are forbidden.