

Version 1.1

Update: 12/27/2019

CERTIFICATE OF ANALYSIS

Product Information

Name	cAMP ELISA Detection Kit
Cat. No.	L00460
Lot No.	2106K014
Component	Anti-IgG Captured Plate Anti-cAMP mAb HRP-cAMP cGMP Standards (0, 1, 3, 9, 27, 81, 243 pmol/ml) cAMP Standard Stock (10 nmol/ml) Assay Buffer A Assay Buffer B 20 × Wash Solution Stop Solution TMB Substrate Plate Sealer User Manual
Storage	The unopened kit is stable for at least 12 months if stored at 2-8 °C and the opened kit is stable for up to 2 weeks at 2-8 °C. Do not freeze the kit.

Quality Control

Test Items	Specifications	Results
Reactivity	Can detect cAMP	Conformed
Lower Detection Limit	0.61 pmol/ml of cAMP	Qualified
Dynamic Range	1-243 pmol/ml	Qualified

Caution:

For laboratory or further manufacturing use only. Not intended for household use. If you have any questions about the Certificate of Analysis, please contact our customer service representative at 1-877-436-7274 (Toll-Free), or 1-732-885-9188.

Certified by: *Dandan Yao*
Quality Department

Date: 06/25/2021

For laboratory research use only. Direct human use, including taking orally and injection and clinical use are forbidden.