

Version 1.1

Update: 12/27/2019

CERTIFICATE OF ANALYSIS

Product Information

| Product Name | GST Tag ELISA Detection Kit | | |
|--------------|---|--|--|
| Cat. No. | L00411 | | |
| Lot No. | 2110K006 | | |
| Kit Content | The kit contains 11 components GST Capture Plate, Sample Dilution Buffer, Biotin-GST | | |
| | | | |
| | Mab Stock, | | |
| | Antibody Dilution Buffer, Streptavidin-HRP, GST | | |
| | Protein Stock (1.0 mg/ml), 20 X Wash Solution, TMB | | |
| | Substrate, Stop Solution, Adhesive Plate Cover, Kit | | |
| | Manual | | |
| Storage | The product is stable for at least one year if stored at | | |
| | 2-8 °C | | |

Quality Control

| Test Items | Specifications | Results | |
|--------------------------|---------------------------------------|-----------|--|
| Reactivity | The kit can detect GST-tagged protein | Passed | |
| Lower Detection Limit | 1ng/ml | Qualified | |
| Dynamic Range | 3.125-100 ng/ml | Qualified | |
| Intra-assay Precision CV | <5% | Qualified | |
| Inter-assay Precision CV | <10% | Qualified | |

Caution:

For laboratory or further manufacturing use only. Not intended for household use. If you have any questions about the Certificate of Analysis, please contact our customer service representative at 1-877-436-7274 (Toll-Free), or 1-732-885-9188.

| Certified by: | Dandan | Yao | Date: 10/22/2021 |
|--------------------|--------|-----|------------------|
| Quality Department | | | |

For laboratory research use only. Direct human use, including taking orally and injection and clinical use are forbidden.