


Sample Description:  
 Structure : HG-31  
 Column : 4.6×250mm, Kromasil C18  
 Solvent A : 0.1% trifluoroacetic in 100% acetonitrile  
 Solvent B : 0.1% trifluoroacetic in 100% water  
 Gradient :  
                   A                  B  
   0.01min      25%              75%  
   25min         55%             45%  
   25.1min     100%              0%  
   30min                          STOP  
 Flow rate : 1.0 mL/min  
 Wavelength : 220nm

Cat. No.: RP12738  
 Lot No.: PJ00509510



### Results

Peak No.	Peak ID	Ret Time	Height	Area	Conc.
1		12.018	60.667	228.757	0.0021
2		13.563	10687.562	86380.867	0.8081
3		14.040	703581.125	10269462.000	96.0734
4		14.040	24544.178	333110.750	3.1163
Total			738873.531	10689182.374	100.0000