

Version :02
Update : 05/08/2023

CERTIFICATE OF ANALYSIS

Product Information

Product Name:	Human IL-8 ELISA Detection Kit
Cat. No.:	L00417
Lot No.:	2303K025
Expiry Data:	28-Mar-2024
Storage condition:	The product is stable for at least one year if stored at 2-8 °C
Kit Content :	IL-8 Capture Plate IL-8 Detection Reagent IL-8 Standard Sample Diluent Assay Diluent 20 × Wash Solution TMB Substrate Stop Solution Adhesive Plate Cover Technical Manual

Quality Control

Test Items	Specifications	Results
Minimum Detectable Dose	<1 ng/ml	Pass
Dynamic Range	25~1600 pg/ml	Pass
R2	>99%	Pass
Zero OD450 Absorbance Value	<0.1 unit	Pass

Date: 05/26/2023

Approved by:  Dandan Yao

Quality Department

For research use only. Not intended for human and animal therapeutic or diagnostic use.

生产商：南京金斯瑞生物科技有限公司 江苏省南京市江宁区科学园雍熙路28号

Manufacturer: Nanjing GenScript Biotech Co., Ltd. No. 28 Yongxi Road, Jiangning District, Nanjing, Jiangsu, China