

Version : 03 Update : 05/08/2023

CERTIFICATE OF ANALYSIS

Product Information

Quality Control	
Expiration data:	14-Aug-2024
Storage condition:	The unopened kit is stable for at least 12 months from the date of manufacture if stored at 2° C to 8° C.
Presentation	96-well plate coated with Goat anti-mouse IgG Anti-cAMP mAb HRP-cAMP Assay Buffer A cAMP Standards Stock cAMP Standard Assay Buffer B 20 × Wash Solution TMB Substrate Stop Solution Plate Sealer User Manual
Kit Content :	Specified on product label
Lot No.:	2308K021
Cat. No.:	L00460
Product Name:	cAMP ELISA Detection Kit

Test Items	Specifications	Results
Intra - Assay	CV≤10%	Pass
B1/B0	75%≤B1/B0≤85%	Pass

Date: 08/21/2023 Approved by: Dandan Yao

Quality Department

For research use only. Not intended for human and animal therapeutic or diagnostic use.

生产商:南京金斯瑞生物科技有限公司 江苏省南京市江宁区科学园雍熙路28号

Manufacturer: Nanjing GenScript Biotech Co., Ltd. No. 28 Yongxi Road, Jiangning District, Nanjing, Jiangsu, China