

Version : 03
Update : 10/30/2023

CERTIFICATE OF ANALYSIS

Product Information

Product Name:	cGMP ELISA Detection Kit
Cat. No.:	L00461
Lot No.:	2309K026
Expiry Date:	25-Sep-2024
Storage condition:	The product is stable for at least one year if stored at 2-8 °C and the opened kit is stable for up to 2 weeks at 2-8 °C. Do not freeze the kit
Kit Content:	Anti-IgG Captured Plate Anti-cGMP pAb HRP-cGMP cGMP Standards (0, 0.3, 0.8, 2.5, 7.4, 22.2, 66.7 pmol/ml) cGMP Standard Stock (10 nmol/ml) Assay Buffer A Assay Buffer B 20 × Wash Solution Stop Solution TMB Substrate Plate Sealer User Manual

Quality Control

Test Items	Specifications	Results
Reactivity	Can detect cGMP	Pass
Lower Detection Limit	0.11 pmol/ml of cGMP	Pass
Dynamic Range	0.3-66.7 pmol/ml	Pass

Date: 11/09/2023

Approved by: *Dandan Yao*

Quality Department

For research use only. Not intended for human and animal therapeutic or diagnostic use.

生产商：南京金斯瑞生物科技有限公司 江苏省南京市江宁区科学园雍熙路28号

Manufacturer: Nanjing GenScript Biotech Co., Ltd. No. 28 Yongxi Road, Jiangning District, Nanjing, Jiangsu, China