

Sample Name : p2Ca
Cat No. : RP19988


Pump A : 0.065% trifluoroacetic in 100% water
Pump A : 0.05% trifluoroacetic in 100% acetonitrile
TotalFlow:1 ml/min
Wavelength:214 nm

Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
25.00	Pumps	Pump A B.Conc	65	
25.10	Pumps	Pump A B.Conc	95	
33.00	Pumps	Pump A B.Conc	95	
33.10	Pumps	Pump A B.Conc	5	
43.00	Pumps	Pump A B.Conc	5	
43.10	Controller	Stop		

<Detector A>

Column : SHIMADZU VP-ODS 4.6 x 150 mm

Chromatogram



1 Det.A Ch1 / 214nm

Peak Table

Detector A Ch1 214nm

Peak#	Ret. Time	Area	Height	Area %
1	20.520	7827695	1076896	96.226
2	22.161	14584	1295	0.179
3	23.803	292412	10133	3.595
Total		8134691	1088324	100.000