

Sample Name : GnRH-I  
 Cat. No. : RP20303  
 Lot. No. : CD10906


Pump A : 0.065% trifluoroacetic in 100% water  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile  
 Total Flow : 1 ml / min  
 Wavelength : 214 nm

Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
25.00	Pumps	Pump A B.Conc	65	
25.10	Pumps	Pump A B.Conc	95	
35.00	Pumps	Pump A B.Conc	95	
35.10	Pumps	Pump A B.Conc	5	
47.00	Pumps	Pump A B.Conc	5	
47.10	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : AlltimaTM C18 4.6 x 250 mm



1 Det.A Ch1 / 214nm

Peak Table

Detector A Ch1 214nm

Peak#	Ret. Time	Area	Height	Area %
1	14.273	19516	2774	0.100
2	14.583	24542	3278	0.126
3	14.907	37254	3096	0.191
4	15.438	96549	15584	0.496
5	15.631	19145955	1967121	98.343
6	15.875	94120	19626	0.483
7	16.313	11316	1488	0.058
8	17.697	25516	1597	0.131
9	20.870	13871	1445	0.071
Total		19468639	2016009	100.000