

Sample Name : OVA Peptide (257-264)
 Cat. No. : RP10611
 Lot. No. : PE5091607

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm


Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
25.00	Pumps	Pump A B.Conc	65	
25.01	Pumps	Pump A B.Conc	95	
31.00	Pumps	Pump A B.Conc	95	
31.01	Pumps	Pump A B.Conc	5	
40.00	Pumps	Pump A B.Conc	5	
40.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : Alltima™ C18 4.6 x 250 mm

Chromatogram



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	16.744	6851	1409	0.276
2	17.211	1068	182	0.043
3	17.620	11099	1549	0.447
4	17.875	13567	2356	0.547
5	18.077	7091	1396	0.286
6	18.250	2416307	374331	97.349
7	18.508	1663	384	0.067
8	18.949	11882	2270	0.479
9	19.193	8335	1557	0.336
10	19.391	1763	319	0.071
11	19.608	2474	483	0.100
Total		2482100	386238	100.000