

## CERTIFICATE OF ANALYSIS

### Product Information

Product Name:	QuickClean II Gel Extraction Kit
Cat. No.:	L00418-100
Lot No.:	C70131804
Exp. Date.:	19 Oct 2019
Storage Condition:	RT

### Kit Contents

Kit Components	100 Assays
Binding Buffer II	85ml
Binding Buffer II	85ml
Wash Buffer	10ml
Wash Buffer	10ml
Elution Buffer	20ml
Spin columns	100(Column Volume:750μl)


### Quality Control

Test Items	Specifications	Results
A260/A280	1.8-2.0	Pass
Electrophoresis	The brightness of the 1000bp PCR band after purified more than 70% that of the band before purified	Pass

### Caution:

For laboratory use only. Not for household use. If you have any questions about the Certificate of Analysis, please contact our customer service representative at 1-877-436-7274 (Toll-Free), or 1-732-885-9188.

Certified by:



Date: 05/08/2018

Product Department

**860 Centennial Ave., Piscataway, NJ 08854, USA**Toll-Free: 1-877-436-7274 Tel: 1-732-885-9188 Fax: 1-732-210-0262 Email: [product@genscript.com](mailto:product@genscript.com) Web: [www.genscript.com](http://www.genscript.com)