

Sample Name :Substance P  
Cat No. : RP10178  
Lot No. : E2571724L  
Time Processed :9:17:49  
Month-Day-Year Processed :12/10/2024

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
Total Flow:1 ml/min  
Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	B.Conc	5
25.00	Pumps	B.Conc	65
25.01	Pumps	B.Conc	95
27.00	Pumps	B.Conc	95
27.01	Pumps	B.Conc	5
35.00	Pumps	B.Conc	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm  
Equipment: ZJ24010222

### <Chromatogram>

mV



### <Peak Table>

Detector A 220nm

Peak#	Ret. Time	Area	Height	Area%
1	9.833	29825	4179	0.363
2	14.658	14359	2779	0.175
3	14.792	8097557	839360	98.627
4	15.200	68550	7458	0.835
Total		8210291	853775	100.000