

Sample Name :Exendin-4  
Cat No. : RP10874  
Lot No. : E2581824L  
Time Processed :11:32:17  
Month-Day-Year Processed :12/18/2024

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
Total Flow:1 ml/min  
Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	B.Conc	5
25.00	Pumps	B.Conc	65
25.01	Pumps	B.Conc	95
27.00	Pumps	B.Conc	95
27.01	Pumps	B.Conc	5
35.00	Pumps	B.Conc	5
35.01	Controller	Stop	

<<Column Performance>>


<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm

Equipment: ZJ23010113

### <Chromatogram>

mV



### <Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	15.625	151699	10884	0.658
2	16.227	77853	4996	0.338
3	16.554	117535	15604	0.510
4	16.927	22620152	1353959	98.140
5	17.271	81706	24505	0.354
Total		23048945	1409948	100.000