

Sample Name :Substance P  
 Cat No. :RP10178  
 Lot No. :E3505321L  
 Time Processed :9:39:31  
 Month-Day-Year Processed :12/31/2021


Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm

Time	Module	Command	Value
0.01	Pumps	B.Conc	5
25.00	Pumps	B.Conc	65
25.01	Pumps	B.Conc	95
27.00	Pumps	B.Conc	95
27.01	Pumps	B.Conc	5
35.00	Pumps	B.Conc	5
35.01	Controller	Stop	

<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm



1 Detector A Channel 1 / 220nm

Peak Table

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	9.316	1356	119	0.065
2	13.842	1092	162	0.052
3	13.975	1881	170	0.090
4	14.730	2039839	182763	97.813
5	15.113	27509	1726	1.319
6	15.580	3310	324	0.159
7	15.779	6896	339	0.331
8	16.491	3559	168	0.171
Total		2085442	185771	100.000