

Sample Name :Substance P
 Cat No. :RP10178
 Lot No. :E6643522D
 Time Processed :14:01:51
 Month-Day-Year Processed :04/25/2022

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>


Time	Module	Command	Value
0.01	Pumps	B.Conc	5
25.00	Pumps	B.Conc	65
25.01	Pumps	B.Conc	95
27.00	Pumps	B.Conc	95
27.01	Pumps	B.Conc	5
35.00	Pumps	B.Conc	5
35.01	Controller	Stop	

<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm
 Equipment: ZJ19010141

<Chromatogram>



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	15.731	12393	841	0.209
2	16.058	36623	2279	0.619
3	16.540	19410	1353	0.328
4	16.975	17385	1668	0.294
5	17.220	41891	3914	0.708
6	17.429	51098	7344	0.864
7	17.488	29500	8936	0.499
8	17.658	5639583	373702	95.329
9	18.013	29634	7180	0.501
10	18.088	38419	4993	0.649
Total		5915936	412208	100.000