

Sample Name :β-Endorphin, human  
Cat No. : RP11344  
Lot No. : E9611624J

Time Processed :21:22:51  
Month-Day-Year Processed :10/18/2024

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
Total Flow:1 ml/min  
Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	B.Conc	5
25.00	Pumps	B.Conc	65
25.01	Pumps	B.Conc	95
27.00	Pumps	B.Conc	95
27.01	Pumps	B.Conc	5
35.00	Pumps	B.Conc	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-SP 4.6 x 250 mm  
Equipment: ZJ19010140

<Chromatogram>

mV



<Peak Table>

Detector A 220nm

Peak#	Ret. Time	Area	Height	Area%
1	16.333	6226915	614795	98.933
2	16.625	48545	6447	0.771
3	16.825	18585	2181	0.295
Total		6294045	623424	100.000