

## CERTIFICATE OF ANALYSIS

### Product Information

---

Product Name	PEGylated Molecule Assay Kit
Cat. No.	L00458
Lot No.	
Kit Content	The kit contains 11 components PEG Molecule Plate, Assay Diluent, Biotin-PEG mAb Stock Antibody Dilution Buffer, Streptavidin-HRP, PEG Standard Control, 20 × Wash Solution, TMB Substrate, Stop Solution, Plate Sealer And User Manual
Storage	The product is stable for at least one year if stored at 2-8 °C

---

### Quality Control

---

Test Items	Specifications	Results
IC <sub>50</sub>	<1 ng/ml for PEG40 kDa	Passed
Range	0.333~81 ng/ml	Passed

---

### Caution

For laboratory or further manufacturing use only. Not intended for household use. If you have any questions about the Certificate of Analysis, please contact our customer service representative at 1-877-436-7274 (Toll-Free), or 1-732-885-9188.

Certified by:   
Antibody Production Director      Date: