

CERTIFICATE OF ANALYSIS

Product Information

Name	cGMP ELISA Kit
Cat. No.	L00461
Lot No.	
Component	Anti-IgG Captured Plate Anti-cGMP pAb HRP-cGMP cGMP Standards (0, 0.3, 0.8, 2.5, 7.4, 22.2, 66.7 pmol/ml) cGMP Standard Stock (10 nmol/ml) Assay Buffer A Assay Buffer B 20 × Wash Solution Stop Solution TMB Substrate Plate Sealer User Manual
Storage	The unopened kit is stable for at least 12 months if stored at 2-8 °C and the opened kit is stable for up to 2 weeks at 2-8 °C. Do not freeze the kit.

Quality Control

Test Items	Specifications	Results
Reactivity	Can detect cGMP	Conformed
Lower Detection Limit	0.11 pmol/ml of cGMP	
Dynamic Range	0.3-66.7 pmol/ml	

Caution

For laboratory or further manufacturing use only. Not intended for household use. If you have any questions about the Certificate of Analysis, please contact our customer service representative at 1-877-436-7274 (Toll-Free), 1-732-885-9188.

Certified by:  Date:

Antibody Production Director