

Sample Name :TAT peptide
 Cat. No. : RP20256
 Lot. No. : PE8351905
 Time Processed :9:02:51
 Month-Day-Year Processed : 05/08/2019

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
25.00	Pumps	Pump A B.Conc	65	
25.01	Pumps	Pump A B.Conc	95	
27.00	Pumps	Pump A B.Conc	95	
27.01	Pumps	Pump A B.Conc	5	
35.00	Pumps	Pump A B.Conc	5	
35.00	Controller	Stop		


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm

Equipment: SS-CM-0286

Chromatogram



1 Det.A Ch1 / 220nm
 Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	7.613	7892	860	0.258
2	8.013	3025475	401080	99.031
3	8.445	17036	3626	0.558
4	9.101	2537	326	0.083
5	9.259	2133	330	0.070
Total		3055074	406223	100.000